

**Manufacturer's Name:** Micro-Vu  
**Manufacturer's Address:** 7909 Conde Lane  
Windsor, California 95492-9779  
United States

**Declares under sole responsibility that the product as originally delivered**

**Product Description:** Multi-Sensor Measuring Machine  
**Product Model(s):** Excel 511, Excel 512, Excel 701, Excel 702, Excel 704, Excel 901,  
Excel 902, Excel 904, Excel 1301, Excel 1302, Excel 1304, Excel 1701,  
Excel 1702, Excel 1704, Excel 2500  
**Product Function:** Dimensional Measurements

**We hereby declare that the equipment listed above conforms to the harmonized standards of the following European Commission Directives:**

- EMC Directive 2014/30/EU
- Low Voltage Directive 2014/35/EU
- Machinery Directive 2006/42/EC
- Rohs Directive 2011/65 / EU


**And conforms with the following product standards:**

	<b>Standard</b>	<b>Limit</b>
<b>EMC</b>	EN 61326-1:2013 EN 55011 Section 6.4: Standard: EN 61000-4-2:2009 Standard: EN 61000-4-3:2010 Standard: EN 61000-4-4:2012 Standard: EN 61000-4-5:2014 Standard: EN 61000-4-6:2014 Standard: EN 61000-4-11:2004	Group 1 Class A Electrostatic Discharge: ±4kV contact, ±8kV air Radiated Immunity: 3V/m Electrical Fast Transients/Burst: ±2kV AC Surges: ±1kV differential mode, ±2kV common mode Conducted Immunity: 3 V, 0.15-80 MHz Supply Dips and Variations: 30% and 100%
<b>Safety</b>	IEC 61010-1:2010 / EN 61010-1:2010 ISO 12100:2010	

**This Declaration is made by:**

02 Aug, 2021

Date



**Ed Amormino**  
Vice President

**Authorized Representative in EC:**

**Name** Ed Amormino  
**Position** General Manager  
Micro-Vu GmbH  
**Address** Essener Strasse 8,  
D-66606 St. Wendel  
Germany  
**Email** Edamormino@microvu.com